Novel Applications of Ultrasound Technology to Visualize and Characterize Myofascial Trigger Points and Surrounding Soft Tissue

Siddhartha Sikdar, PhD, Jay P. Shah, MD, Tadesse Gebreab, BS, Ru-Huey Yen, BS, Elizabeth Gilliams, BS, Jerome Danoff, PT, PhD, Lynn H. Gerber, MD

ABSTRACT. Sikdar S, Shah JP, Gebreab T, Yen R-H, Gilliams E, Danoff J, Gerber LH. Novel applications of ultrasound technology to visualize and characterize myofascial trigger points and surrounding soft tissue. Arch Phys Med Rehabil 2009;90:1829-38.

Objective: To apply ultrasound (US) imaging techniques to better describe the characteristics of myofascial trigger points (MTrPs) and the immediately adjacent soft tissue.

Design: Four sites in each patient were labeled based on physical examination as active myofascial trigger points (A-MTrPs; spontaneously painful), latent myofascial trigger points (L-MTrPs; nonpainful), or normal myofascial tissue. US examination was performed on each subject by a team blinded to the physical findings. A 12~5MHz US transducer was used. Vibration sonoelastography (VSE) was performed by color Doppler variance imaging while simultaneously inducing vibrations (\sim 92Hz) with a handheld massage vibrator. Each site was assigned a tissue imaging score as follows: 0, uniform echogenicity and stiffness; 1, focal hypoechoic region with stiff nodule; 2, multiple hypoechoic regions with stiff nodules. Blood flow in the neighborhood of MTrPs was assessed using Doppler imaging. Each site was assigned a blood flow waveform score as follows: 0, normal arterial flow in muscle; 1, elevated diastolic flow; 2, high-resistance flow waveform with retrograde diastolic flow.

Setting: Biomedical research center.

Participants: Subjects (N=9) meeting Travell and Simons' criteria for MTrPs in a taut band in the upper trapezius.

Interventions: Not applicable.

Main Outcome Measures: MTrPs were evaluated by (1) physical examination, (2) pressure algometry, and (3) three types of US imaging including gray-scale (2-dimensional [2D] US), VSE, and Doppler.

Results: MTrPs appeared as focal, hypoechoic regions on 2D US, indicating local changes in tissue echogenicity, and as

0003-9993/09/9011-00204\$36.00/0

doi:10.1016/j.apmr.2009.04.015

focal regions of reduced vibration amplitude on VSE, indicating a localized, stiff nodule. MTrPs were elliptical, with a size of $.16\pm.11$ cm². There were no significant differences in size between A-MTrPs and L-MTrPs. Sites containing MTrPs were more likely to have a higher tissue imaging score compared with normal myofascial tissue (*P*<.002). Small arteries (or enlarged arterioles) near A-MTrPs showed retrograde flow in diastole, indicating a highly resistive vascular bed. A-MTrP sites were more likely to have a higher blood flow score compared with L-MTrPs (*P*<.021).

Conclusions: Preliminary findings show that, under the conditions of this investigation, US imaging techniques can be used to distinguish myofascial tissue containing MTrPs from normal myofascial tissue (lacking trigger points). US enables visualization and some characterization of MTrPs and adjacent soft tissue.

Key Words: Myofascial pain syndromes; Rehabilitation; Sonoelastography; Trigger points, myofascial; Ultrasonography.

Published by Elsevier Inc on behalf of the American Congress of Rehabilitation Medicine.

C HRONIC PAIN IS A CRITICAL public health problem.¹ A vast number of patients in specialty pain management centers and 95% of people with chronic pain disorders have MPS.² MPS is a common, nonarticular musculoskeletal disorder, characterized by MTrPs—hard, palpable, discrete, localized nodules located within taut bands of skeletal muscle that are painful on compression. MTrPs can be either active or latent.³ An A-MTrP is associated with spontaneous pain in which pain is present without palpation. This spontaneous pain can be at the site of the MTrP or remote from it. However, firm palpation of the A-MTrP increases pain locally and usually reproduces the subject's remote pain.⁴ An L-MTrP is not associated with spontaneous pain, although pain can often be elicited in an asymptomatic subject by a mechanical stimulus,

List of Abbreviations

A-MTrP L-MTrP MPS MRE MTrP Po ₂ PPT RI 3D 2D US	active myofascial trigger point latent myofascial trigger point myofascial pain syndrome magnetic resonance elastography myofascial trigger point partial pressure of oxygen pain pressure threshold resistive index 3-dimensional 2-dimensional ultrasound
US	ultrasound
VOE	vibration sonoelastography

From the Department of Electrical and Computer Engineering (Sikdar) and Center for the Study of Chronic Illness and Disability (Gerber), George Mason University, Fairfax, VA; and Rehabilitation Medicine Department, National Institutes of Health, Bethesda, MD (Shah, Gebreab, Yen, Gilliams, Danoff, Gerber).

Presented in part to the 30th Annual International Conference of the Institute for Electrical and Electronics Engineers, Engineering in Medicine and Biology Society, August 20–24, 2008, Vancouver, BC, Canada, and the 69th Annual Assembly of the American Academy of Physical Medicine and Rehabilitation, November 20–23, 2008, San Diego, CA.

Supported by the Intramural Research Program, National Institutes of Health (NIH), and the Clinical Center and Office of the Director, NIH.

No commercial party having a direct financial interest in the results of the research supporting this article has or will confer a benefit on the authors or on any organization with which the authors are associated.

Correspondence to Siddhartha Sikdar, PhD, Department of Electrical and Computer Engineering, George Mason University, 4400 University Dr, MS 1G5, Fairfax, VA 22030, e-mail: *ssikdar@gmu.edu*.

such as finger pressure over the L-MTrP.⁵ In someone with a spontaneous pain complaint, thorough palpation of the myofascial tissue is required to identify and differentiate an A-MTrP from an L-MTrP. Pain elicited by palpation of an L-MTrP in a symptomatic subject is qualitatively different from the subject's pain complaint.

Despite the high prevalence of MPS,^{1,2,6-9} its pathophysiology is unclear. It is unknown whether the nodules are a result of ongoing anatomic and physiologic abnormalities, or whether they occur independently of other abnormalities in surrounding soft tissue. The current diagnostic standard for myofascial pain is based on palpation for the presence of trigger points in a taut band of skeletal muscle and an associated symptom cluster that includes referred pain patterns.³ Unfortunately, few physicians receive training in the clinical diagnosis of myofascial pain. Moreover, the physical examination has been reported to be unreliable,¹⁰ and until recently, there has been no clearly demonstrable underlying pathology associated with the physical findings of trigger points and taut bands.¹¹⁻¹⁵

Current approaches for pain relief include needling (with or without injection) and massage therapy. The lack of objective clinical outcome measures has been a barrier for critically evaluating the efficacy of these therapeutic methods. All of these factors have led to a lack of consensus on myofascial pain as a clinical entity and have contributed to the uncertainty about the pathogenesis and pathophysiology of trigger points.¹⁶ Therefore, there is a need to develop objective, repeatable, and reliable diagnostic tests for identifying MTrPs and evaluating treatment outcome measures. Such measures can be used to properly diagnose and understand the natural history of MTrPs and to determine the underlying mechanisms relevant to the development, amplification and resolution of myofascial pain.

Diagnostic US is used extensively for noninvasive real-time imaging of muscle, tendon, fascia, blood vessels, and other soft tissues and is suitable for use in a medical office setting. US has the potential to characterize viscoelastic properties of myofascial tissue,¹⁵ to quantify hemodynamic changes resulting from compression of blood vessels, to provide dynamic measures of tissue performance (eg, muscle contraction), and to demonstrate structure/function correlations.¹⁷⁻¹⁹ US is a low-risk method for obtaining descriptive information of tissue (eg, presence of fat, fiber, fluid) and its mechanical properties.^{20,21} To our knowledge, Doppler US, which is routinely used clinically for vascular applications, has not been used to study MTrPs. However, VSE is a validated method to image palpably stiff nodules in tissue.²² When an external vibration is applied, any localized regions of stiffer tissue vibrate with lower amplitude compared with less stiff surrounding tissue.²³ In VSE, the US color variance mode can be used to image the relative distribution of vibration amplitude in the myofascial tissue and identify any localized regions of stiffness. Several studies have demonstrated the feasibility of VSE. Transient VSE has been used to quantify the anisotropic properties of muscle and to measure the shear elastic moduli of relaxed and contracted states of both the quadriceps²⁴ and biceps brachii²⁵ muscles. A different imaging technique, magnetic resonance elastography, has also been used for investigating taut bands.¹³

In this study, we investigated the feasibility of US imaging for visualizing trigger points and surrounding soft tissue. We believe that US imaging techniques will be capable of distinguishing MTrPs and adjacent structures from normal myofascial tissue and will provide objective descriptions of the underlying tissue abnormalities associated with MTrPs, which may help decipher the pathophysiology and the pathogenesis of painful MTrPs.

METHODS

Study Population

This descriptive study was carried out at the Rehabilitation Medicine Department of the National Institutes of Health, Clinical Research Center. Subjects with acute cervical pain (<3mo) were eligible and met inclusion criteria if found to have an A-MTrP in 1 or both upper trapezii. All subjects underwent a thorough musculoskeletal evaluation to rule out potential causes of their symptoms other than MTrPs. Exclusion criteria included muscle pain due to fibromyalgia and atypical facial neuralgia, a history of myopathy, neck and shoulder conditions including cervical radiculopathy and lumbosacral myopathy, a history of cervical spine or shoulder surgery, and a history of trigger point injections in the upper trapezius. In addition, subjects with cancer or with infections of the head, eyes, ears, nose, or throat were excluded from the study. Based on their history and physical findings in the upper trapezius muscles, 9 subjects with a diagnosis of myofascial pain were recruited. All but 2 had unilateral pain symptoms. The Institutional Review Board of the National Institute of Dental and Craniofacial Research approved this study, and each participant provided informed consent to participate in the study.

Clinical Examination

Subjects underwent a physical examination by an experienced physiatrist (J.P.S.), who determined the presence or absence of MTrPs in the upper trapezius muscle according to the standard clinical criteria defined by Travell and Simons.³ In this study, any local region of myofascial tissue in which nodules were absent to palpation was defined as normal or uninvolved.

In our study design, we sought to identify 2 sites per side on each subject including a variety of A-MTrPs, L-MTrPs, and normal sites. Each subject was to have at least 1 normal site among the 4. Palpation was in the central region of the upper trapezius muscle within 6cm of the muscle's midline (approximately midway between the cervical vertebrae and the acromion process). Up to 2 nodules were found by the examiner in each upper trapezius and identified as either an A-MTrP or an L-MTrP. If fewer than 2 nodules were identified, palpation continued until the examiner was satisfied that only 1 nodule or none was present in the muscle. The examiner then marked "dummy" sites in the same general vicinity as the nodules and recorded them as "normal." This process resulted in 2 marked sites on each upper trapezius (4 total per subject). Pressure algometry^a was performed on each of the 4 sites to determine the PPT (ie, the amount of pressure that produces the sensation of pain). Only the examiner knew the clinical status and classifications of the marked sites (ie, the sonography team [S.S., T.G.] was blinded to the clinical status, algometry results, and identity of all sites). Approximately 30 minutes elapsed between PPT procedures and sonography.

Gray-Scale Imaging

Each participant underwent a US examination using a Philips $iU22^{b}$ clinical US system with a $12\sim5MHz$ linear array L12-5 transducer targeted at the 4 sites palpated during the clinical examination. An experienced sonographer (with >18y of experience) performed all US examinations. The upper trapezius was visualized in the longitudinal and transverse views with the subject sitting upright in a comfortable position. On 2D gray-scale imaging, MTrPs in the upper trapezius appeared as focal hypoechoic (darker) areas with heteroge-



Fig 1. Design for a vibration source that can be used for VSE imaging of the upper trapezius. This design can induce vibrations uniformly over a broad area both along the muscle fibers (A) and transverse to the muscle fibers (B).

neous echotexture. Because a single 2D image did not fully capture the 3D area of the trigger point and surrounding soft tissue, 3D imaging was also performed by manually sweeping the transducer over the upper trapezius along the transverse and longitudinal planes. A mechanically scanned $9\sim3MHz$ 3D transducer (3D9-3v) was used for real-time 3D imaging of trigger points in the upper trapezius of the subjects.

Elastography Imaging

Ultrasound VSE uses an external vibration source less than 1000Hz in conjunction with Doppler techniques to identify localized regions of increased tissue stiffness. This methodology has been described in detail elsewhere.²² In this study, vibrations were induced in the upper trapezius muscle using an external handheld vibrating massager (model NC70209^c) modified with a flat attachment head (fig 1). This vibration source was placed approximately 2 to 3cm from each of the 4 sites to be imaged. Color variance mode was then used to image the site while the massager induced vibrations of approximately 92Hz in the muscle.

Doppler Imaging

We investigated the vasculature and circulation in and around the marked sites using Doppler US. The peak systolic velocity and the minimum diastolic velocity blood flow velocities were measured in the ascending branch of the transverse cervical artery and any other arteries or enlarged arterioles that were found in the neighborhood of palpable trigger points. The RI is commonly used for vascular diagnosis and is defined as the ratio of the difference in peak systolic and minimum diastolic velocities to the peak systolic velocity. The RI is an indication of the resistance of the end-organ vascular bed. In muscle, normally RI equals 1, indicating no diastolic flow. Elevated diastolic flow (RI<1) indicates decreased vascular bed resistance, while negative diastolic flow (RI>1) indicates increased vascular bed resistance.

Ordinal Imaging Score

Two ordinal scores were devised to describe tissue and blood flow characteristics of the imaged sites. The tissue imaging score, based on gray-scale and elastography imaging, was given a range from 0 (normal, uniform echogenicity and stiffness) to 2 (abnormal structure with multiple focal hypoechoic and stiff nodules). The blood flow waveform score, based on the Doppler flow waveform, was given a range from 0 (normal arterial flow) to 2 (abnormal high-resistance flow with retrograde diastolic flow). Table 1 lists the scoring criteria.

Two investigators (T.G., S.S.) independently quantified the size of the MTrPs from the acquired US images. An ellipse was manually fitted to the most prominent MTrP, and the area was computed using the measurement tool available on the US scanner.

Table 1: Ordinal Imaging Scores

Tissue Imaging Score		Blood Flow Waveform Score	
Score	Criterion	Score	Criterion
0	No focal lesion on echo or stiffness image (includes heterogeneity)	0	Normal muscle flow or no visible blood vessel (RI*=1)
1	Evidence of focal lesion on both echo and stiffness image	1	Elevated blood flow in diastole (RI*<1)
2	Multiple focal lesions or marked heterogeneity on both echo and stiffness image	2	Oscillatory flow or sustained retrograde flow in diastole (RI*>1)

*RI, the ratio of the difference in peak systolic and minimum diastolic velocities to the peak systolic velocity.



Fig 2. PPTs (in lb) measured using pressure algometry demonstrated lower thresholds in A-MTrPs and L-MTrPs compared with palpably normal muscle (N>A, L; P<.007). The error bars correspond to SDs.

Statistical Analysis

PPTs were compared among sites containing A-MTrPs, L-MTrPs, and normal myofascial tissue using t tests subject to Bonferroni correction. The nonparametric Kruskal-Wallis test was used to compare the tissue imaging score and the blood flow score among sites containing A-MTrPs, L-MTrPs, and normal myofascial tissue. Pearson's correlation coefficient was used to assess the interrater reliability of the MTrP size measurement. For all tests, a P value less than .05 was considered significant.

RESULTS

The age range of the 9 subjects (2 men, 7 women) was 23 to 55 years. US data were acquired from 33 sites in the upper trapezius of the 9 subjects (2 sites in 1 subject and 1 site in another subject could not be imaged because of scheduling restrictions). Based on the physical examination, 13 sites were classified as A-MTrPs, 6 sites were classified as L-MTrPs, and 14 sites were classified as normal. Eleven of the 13 A-MTrPs, 5 of the 6 L-MTrPs, and 1 of the 14 normal sites were found at a more medial location on the upper trapezius (relative to the midline of the body), while the rest were found at a more lateral location. Pressure algometry was performed at 30 of these sites (algometry was not performed in 1 patient).

PPTs determined using pressure algometry showed that sites with A-MTrPs and L-MTrPs had a significantly lower threshold compared with palpably normal sites (P<.01) (fig 2). The PPTs were not significantly different between A-MTrPs and L-MTrPs.

On US imaging, MTrPs in the upper trapezius muscle appeared as elliptically shaped focal areas of hypoechogenicity that corresponded with the location of the palpable nodule in all the subjects examined (fig 3A). In our images, trigger points often appear to coexist with multiple nodules in close proximity (fig 3B). The size of MTrPs was $.16\pm.11$ cm² (mean \pm SD for both examiners and for all subjects). No significant size difference was found between A-MTrPs ($.16\pm.11$ cm²) and L-MTrPs ($.15\pm.13$ cm²). When quantifying size, there was good agreement between the examiners (Pearson's r=.78, P<.001, excluding 1 outlying case where the first examiner identified a single large MTrP, while the second examiner identified 2 smaller MTrPs).

In 3D imaging, the discrete trigger point was clearly visible in the longitudinal, transverse, and coronal images (fig 4). The echotexture in the nodule was heterogeneous and markedly different from a control area in the upper trapezius muscle found to be palpably normal on physical examination. The palpably normal tissue appeared isoechoic with homogeneous echotexture (compare fig 5A and 5C).

On color variance imaging, MTrPs appeared as focal areas of reduced vibration amplitude. Figure 5D shows the color variance image of the upper trapezius muscle with a palpable MTrP indicated by an arrow. The area of reduced vibration amplitude corresponds with the hypoechoic regions observed on 2D gray-scale imaging for all subjects. Figure 5B shows the color variance image of the upper trapezius muscle that was normal on physical examination. The entire region of the muscle appears to vibrate with approximately uniform amplitude, as indicated by the uniform color. In some regions, single, well-defined nodules were apparent, while in others multiple focal nodules could be observed (fig 5E, 5F).

Color Doppler and duplex Doppler examination of the upper trapezius in our subject sample revealed differences between A-MTrPs and L-MTrPs. Prominent blood vessels were found at 27 of the 33 sites. In some cases, the blood vessel passed directly through a trigger point (figs 6 and 7). In 9 (69%) of 13 sites with A-MTrPs, we observed unique Doppler flow waveforms (fig 8) with retrograde diastolic flow. Intriguingly, retrograde diastolic flow was observed in only 1 (16.7%) of 6 sites with an L-MTrP. Only 1 palpably normal site of 14 (7%) showed sustained retrograde diastolic flow. This site was adjacent to a site with an A-MTrP.



Fig 3. Gray-scale imaging of a trigger point in the upper trapezius. (A) An isolated MTrP appears as a well-defined focal hypoechoic nodule. (B) A series of 4 hypoechoic MTrPs in the upper trapezius.



Fig 4. 3D imaging of trigger points. A mechanically scanned 3D probe (3D9-3v) was used for 3D imaging in a subject with a latent trigger point. The MTrP is clearly identified (arrows) in all 3 planes as well as in a multislice view.

To quantify these findings, we used a composite ordinal scoring system using the tissue imaging score and blood flow score as variables (see table 1). The number of sites and their scores based on US imaging for active, latent, and normal sites in the upper trapezius muscle are shown in figure 9. Palpable trigger points identified on physical examination had a significantly higher tissue imaging score compared with normal myofascial tissue ($P \le .002$ using the nonparametric Kruskal-Wallis test), indicating that the imaging findings correlated with the physical findings. Tissue imaging scores were not significantly different between A-MTrPs and L-MTrPs. However, A-MTrPs had a significantly higher blood flow waveform score compared with L-MTrPs in the upper trapezius (P < .021, Kruskal-Wallis), indicating that A-MTrPs are associated with blood flow abnormalities. Blood flow score was not significantly different between L-MTrPs and normal sites.

DISCUSSION

This study was undertaken to determine whether a noninvasive, easily accessible and relatively inexpensive technology could be used to characterize MTrPs. Our preliminary findings indicate that palpable nodules (identified as either A-MTrPs or L-MTrPs) appear as 1 or more focal nodules on gray-scale and color variance US imaging and are absent in palpably normal myofascial tissue. Some hypoechoic regions were observed even when nodules were not palpable. We did not investigate the ultrasonic characteristics of taut bands in this study. A-MTrPs and L-MTrPs demonstrated different and distinct blood flow waveform patterns. Significantly more A-MTrPs than L-MTrPs were associated with retrograde flow in diastole, which would be indicative of a highly resistive vascular bed.

A-MTrPs have 2 clinical attributes that must be investigated in order to elucidate the pathogenesis and pathophysiology of myofascial pain. One attribute is a motor dysfunction of the myofascial tissue as indicated by the finding of a constant, discrete hardness, usually palpable as a nodule in a taut band within the belly of the muscle. The other attribute is a sensory abnormality characterized primarily by tenderness and spontaneous pain. The pain can be localized to the immediate area of the nodularity/taut band or be referred to a distant part of the body, or both.

Several promising lines of scientific study (ie, histologic, neurophysiologic, biochemical, and somatosensory) have attempted to explain the pathophysiologic basis of the motor and sensory abnormalities associated with MTrPs.^{11,12,26-30} Findings from these studies support Simons' Integrated Hypothesis, which postulates that an "energy crisis" perpetuates an initial sustained sarcomere contracture, which leads to increased local metabolic demands in the presence of compromised capillary circulation. This leads to local hypoxia, tissue damage, or both. Shah et al^{11,12} found and confirmed the presence of elevated levels of inflammatory mediators, catecholamines, neuropeptides, and proinflammatory cytokines in the vicinity of A-MTrPs. These biochemicals are known to be associated with persistent pain states, myofascial tenderness, intercellular signaling, and inflammation. These attributes may help explain the sensory abnormalities associated with A-MTrPs.3

Results of 2D US imaging confirm that significant tissue abnormalities and morphological changes are associated with MTrPs. Differences in echogenicity and stiffness of the MTrP compared with the surrounding tissue suggest a disruption of normal muscle fiber structure and a change in local tissue characteristics. The hypoechoic and stiffer nodules may be indicative of contraction knots resulting from increased muscle fiber contraction and recruitment, local injury, and/or localized regions of ischemia.



Fig 5. Simultaneous 2D grayscale and color variance imaging. (A and B) Normal upper trapezius muscle. The normal muscle appears isoechoic and has uniform color variance (TIS=0). (C and D) Muscle with a palpable MTrP. A hypoechoic region and a well-defined focal decrease of color variance indicating a localized stiffer region is visible (TIS=1). (E and F) Muscle with a palpable MTrP. Multiple hypoechoic regions and multiple focal nodules are visible (TIS=2). Abbreviation: TIS, tissue imaging score.



Fig 6. 3D color Doppler imaging of blood vessels passing through trigger points. A mechanically scanned 3D probe (3D9-3v) was used for 3D imaging in a subject with a latent trigger point. The blood vessel is clearly visualized in all 3 planes.

To date, we know of only one other study that has attempted to objectively characterize the physical features of MTrPs. In this study by Chen et al,³² MRE was used to measure the viscoelastic properties of the upper trapezius muscles. MRE uses a modified gradient echo pulse sequence to image the propagation of induced vibration shear waves. Although they were able to demonstrate that the shear wave propagation pattern in a taut band differs from that in palpably normal myofascial tissue, the MTrP itself was not identified within the taut band. US imaging may be a better method to localize the MTrP while simultaneously investigating the neighboring tissue structure and vasculature. 2D US combined with VSE provides both visual imaging and objective measurements of the physical characteristics and mechanical properties of the MTrP. We believe that this novel application of imaging US will provide significant methodological improvement over MRE, enabling costeffective imaging and longitudinal monitoring of MTrPs in an office-based setting. Using a transient vibration source, the velocity of the propagating shear wave could be measured using US, which would allow further quantification of the tissue stiffness.²⁴ US provides new and exciting possibilities for identifying physical characteristics of MTrPs on human subjects in vivo, noninvasively and at low cost.

Another advantage of US imaging is the ability to investigate the vasculature and blood flow characteristics in and around MTrPs. Results of our blood flow investigation associate blood flow disturbances with the pathophysiology of A-MTrPs. Doppler flow waveforms of arteries in the neighborhood of A-MTrPs showed high-resistance blood flow with retrograde diastolic flow, which differed from the blood flow around L-MTrPs and normal, uninvolved myofascial tissue. An increase in vascular resistance in A-MTrPs is consistent with blood vessel compression caused by sustained contracture at a trigger point or constriction caused by oxidative stress or hypoxia. The blood vessel compression may be sufficient to cause the local hypoperfusion, hypoxia, and other pathophysiologic developments leading to the pain, tenderness, and nodularity of an A-MTrP.

Jarvholm et al³³ found that high intramuscular pressure in the supraspinatus muscle significantly impeded local muscle blood flow, which would lead to hypoperfusion and local ischemia. Measurements of the Po2 in myogeloses (which presumably are accumulations of several MTrPs) showed that the Po₂ is extremely low (close to 0) at the center of an MTrP area.³⁴ The low Po₂ is likely to be associated with a lack of adenosine triphosphate, a condition that may contribute to the local hypercontractions (contractures) found in this area. Adenosine triphosphate is needed to break the bonds between the myofilaments and to end the muscle contraction. The available data on myofascial pain suggest that, in fact, pain associated with A-MTrPs may be caused indirectly by decreased blood flow to the MTrP. The hypoperfusion may lead to the observed low Po_2 and the contractures. A low Po_2 is known to be a powerful factor for the release of bradykinin, a sensitizing agent for muscle nociceptors.

The US techniques presented in this article can be used to identify anatomic and physiologic abnormalities associated with MTrPs. These findings will help to establish objective diagnostic criteria for the identification and classification of MTrPs, which would be more reliable, sensitive, and specific than physical examination alone. Changes in these objective measures in response to treatment, such as alterations in the blood flow or changes in the size, appearance, or stiffness of trigger points, or both, could then be used as treatment outcome measures. Further studies are needed to confirm these findings and their clinical usefulness.



Fig 7. (A) The main blood supply to the upper trapezius is through the ascending branch of the transverse cervical artery. Unless stated otherwise, this image is from the 20th U.S. edition of Gray's Anatomy of the Human Body, originally published in 1918 and therefore lapsed into the public domain. (B) The ascending branch can be visualized using color Doppler imaging. The blood flow waveform in the ascending branch or other branches arising from this vessel can provide an indication of the flow resistance in the perfused tissue. (C) A blood vessel passing through an A-MTrP.

Study Limitations

We believe that our observations of echogenicity, stiffness, and blood flow may enable differentiating A-MTrPs from L-MTrPs through US visualization. However, this study is exploratory and descriptive, and the findings are from a small number of subjects. Therefore, universal generalization to MPS and MTrPs is premature, as is development of a definitive model to explain etiology of MTrPs. In addition, there are some technical difficulties, which we are currently addressing. For example, the operator technique (amount of pressure and the angle at which the US transducer head is held to scan the tissue) is difficult to control. We are attempting to standardize the scan technique for future investigations. The vibrator used in our study for qualitative sonoelastography imaging was a modified off-the-shelf handheld device; for more quantitative studies improved designs are



Fig 8. (A) Subject with an A-MTrP visible as a hypoechoic region on the gray-scale image, and an artery running through the MTrP visible on color Doppler (the Doppler sample volume is placed inside the MTrP). High-resistance blood flow waveform with retrograde diastolic flow was observed in the artery. (B) The same subject had an L-MTrP on the contralateral side with an artery running through it, which showed elevated diastolic flow but no retrograde diastolic flow. (C) Four waveform shapes observed in our studies. C.1 shows arterial flow in muscle with no diastolic flow (BFS=0). C.2 shows elevated flow in diastole (BFS=1). C.3 shows oscillatory high-resistance flow with retrograde flow in early diastole (BFS=2). C.4 shows sustained retrograde flow in diastole (BFS=2). Abbreviation: BFS, blood flow score.



Fig 9. Distribution of scores based on US imaging for (A) active, (B) latent, and (C) normal sites in upper trapezius muscle. Muscle with palpable trigger points on clinical exam (either active or latent) had a significantly higher TIS compared with palpably normal muscle (P<.002). One-hundred percent of A-MTrPs and L-MTrPs had a TIS of 1 or 2, compared with 36% of normal sites. A-MTrPs had a significantly higher BFS compared with L-MTrPs (P<.029). Sixty-nine percent of active sites had a BFS of 2 compared with 16.7% of latent sites. Abbreviations: BFS, blood flow score; TIS, tissue imaging score.

possible. While our methodology was designed to keep the sonography team blinded to the clinical findings, unintentional bias cannot be ruled out because the nodules were readily visualized on imaging. Our study did not include a control group of pain-free subjects. All but 2 subjects had unilateral symptoms, and the contralateral side in the same subject was used as the control. However, some bias resulting from the anatomic location of the MTrPs cannot be ruled out. Finally, interobserver and intraobserver variabilities were not assessed in this feasibility study. Nonetheless, we believe that the reported findings expand the understanding of MTrPs and bring us closer towards a model of the pathophysiology of the MTrPs and their surrounding milieu.

CONCLUSIONS

Under the conditions of this investigation, gray-scale echogenicity and color variance imaging based on relative stiffness were used to differentiate palpable nodules in soft tissue (either active or latent) from normal myofascial tissue, and blood flow waveform characteristics were used to differentiate A-MTrPs and L-MTrPs. Our preliminary findings show that US, a convenient, accessible, and low-risk technique, can be a useful method for differentiating MTrPs from surrounding tissue.

References

- Gerwin RD. Classification, epidemiology, and natural history of myofascial pain syndrome. Curr Pain Headache Rep 2001; 5:412-20.
- Fishbain DA, Goldberg M, Meagher BR, Steele R, Rosomoff H. Male and female chronic pain patients categorized by DSM-III psychiatric diagnostic criteria. Pain 1986;26:181-97.
- Simons DG, Travell JG, Simons PT. Travell and Simons' myofascial pain and dysfunction: the trigger point manual. I. Upper half of body. 2nd ed. Baltimore: Williams and Wilkins; 1999.
- Fernandez-de-Las-Penas C, Simons D, Cuadrado ML, Pareja J. The role of myofascial trigger points in musculoskeletal pain syndromes of the head and neck. Curr Pain Headache Rep 2007; 11:365-72.
- Lucas KR. The impact of latent trigger points on regional muscle function. Curr Pain Headache Rep 2008;12:344-9.
- Fernandez-de-Las-Penas C, Alonso-Blanco C, Cuadrado ML, Gerwin RD, Pareja JA. Myofascial trigger points and their relationship to headache clinical parameters in chronic tension-type headache. Headache 2006;46:1264-72.

- Rosomoff HL, Fishbain DA, Goldberg M, Santana R, Rosomoff RS. Physical findings in patients with chronic intractable benign pain of the neck and/or back. Pain 1989;37:279-87.
- Calandre EP, Hidalgo J, Garcia-Leiva JM, Rico-Villademoros F. Trigger point evaluation in migraine patients: an indication of peripheral sensitization linked to migraine predisposition? Eur J Neurol 2006;13:244-9.
- Skootsky SA, Jaeger B, Oye RK. Prevalence of myofascial pain in general internal medicine practice. West J Med 1989;151:157-60.
- Hsieh CY, Hong CZ, Adams AH, et al. Interexaminer reliability of the palpation of trigger points in the trunk and lower limb muscles. Arch Phys Med Rehabil 2000;81:258-64.
- Shah JP, Danoff JV, Desai MJ, et al. Biochemicals associated with pain and inflammation are elevated in sites near to and remote from active myofascial trigger points. Arch Phys Med Rehabil 2008;89:16-23.
- Shah JP, Phillips TM, Danoff JV, Gerber LH. An in vivo microanalytical technique for measuring the local biochemical milieu of human skeletal muscle. J Appl Physiol 2005;99:1977-84.
- Chen Q, Bensamoun S, Basford JR, Thompson JM, An KN. Identification and quantification of myofascial taut bands with magnetic resonance elastography. Arch Phys Med Rehabil 2007; 88:1658-61.
- Simons DG. New views of myofascial trigger points: etiology and diagnosis. Arch Phys Med Rehabil 2008;89:157-9.
- Sikdar S, Shah JP, Gilliams E, Gebreab T, Gerber LH. Assessment of myofascial trigger points (MTrPs): a new application of ultrasound imaging and vibration sonoelastography. Conf Proc IEEE Eng Med Biol Soc 2008;1:5585-9.
- Wheeler AH. Myofascial pain disorders: theory to therapy. Drugs 2004;64:45-62.
- Chi-Fishman G, Hicks JE, Cintas HM, Sonies BC, Gerber LH. Ultrasound imaging distinguishes between normal and weak muscle. Arch Phys Med Rehabil 2004;85:980-6.
- Hicks JE, Hawker TH, Jones BL, Linzer RT, Gerber LH. Diagnostic ultrasound: its use in the evaluation of muscle. Arch Phys Med Rehabil 1984;65:129-31.
- Saltzstein RJ, Anoff JV, Shawker TH, Jones BL, Gerber L. Ultrasound sector scanning used to define changes in muscle configuration. Scand J Rehabil Med 1989;21:209-12.
- Sikdar S, Beach KW, Vaezy S, Kim Y. Ultrasonic technique for imaging tissue vibrations: preliminary results. Ultrasound Med Biol 2005;31:221-32.

- Shamdasani V, Bae U, Sikdar S, et al. Research interface on a programmable ultrasound scanner. Ultrasonics 2008;48:159-68.
- Taylor LS, Porter BC, Rubens DJ, Parker KJ. Three-dimensional sonoelastography: principles and practices. Phys Med Biol 2000; 45:1477-94.
- Gao L, Parker KJ, Alam SK, Lernel RM. Sonoelasticity imaging: theory and experimental verification. J Acoust Soc Am 1995;97: 3875-86.
- Levinson SF, Shinagawa M, Sato T. Sonoelastic determination of human skeletal muscle elasticity. J Biomech 1995;28:1145-54.
- Gennisson JL, Cornu C, Catheline S, Fink M, Portero P. Human muscle hardness assessment during incremental isometric contraction using transient elastography. J Biomech 2005;38:1543-50.
- Mense S, Simons DG, Hoheisel U, Quenzer B. Lesions of rat skeletal muscle after local block of acetylcholinesterase and neuromuscular stimulation. J Appl Physiol 2003;94:2494-501.
- Reitinger A, Radner H, Tilscher H, Hanna M, Windisch A, Feigl W. Morphologische untersuchung an triggerpunkten. Manuelle Medizin 1996;34:256-62.
- 28. Windisch A, Reitinger A, Traxler H, et al. Morphology and histochemistry of myogelosis. Clin Anat 1999;12:266-71.
- 29. Kuan TS, Hsieh YL, Chen SM, Chen JT, Yen WC, Hong CZ. The myofascial trigger point region: correlation between the degree of irritability and the prevalence of endplate noise. Am J Phys Med Rehabil 2007;86:183-9.

- Niddam DM, Chan RC, Lee SH, Yeh TC, Hsieh JC. Central modulation of pain evoked from myofascial trigger point. Clin J Pain 2007;23:440-8.
- 31. Shah JP, Gilliams EA. Uncovering the biochemical milieu of myofascial trigger points using in vivo microdialysis: an application of muscle pain concepts to myofascial pain syndrome. J Bodyw Mov Ther 2008;12:371-84.
- Chen Q, Basford J, An KN. Ability of magnetic resonance elastography to assess taut bands. Clin Biomech (Bristol, Avon) 2008;23:623-9.
- Jarvholm U, Styf J, Suurkula M, Herberts P. Intramuscular pressure and muscle blood flow in supraspinatus. Eur J Appl Physiol Occup Physiol 1988;58:219-24.
- 34. Strobel ES, Krapf M, Suckfull M, Bruckle W, Fleckenstein W, Muller W. Tissue oxygen measurement and 31P magnetic resonance spectroscopy in patients with muscle tension and fibromyalgia. Rheumatol Int 1997;16:175-80.

Suppliers

- Pain Diagnostics and Thermography, 233 E Shore Rd, Ste 108, Great Neck, NY 11023-2433.
- b. Philips Healthcare USA, 22100 Bothell Everett Hwy, Bothell, WA 98021-8431.
- c. North Coast Medical, Inc, 18305 Sutter Blvd, Morgan Hill, CA 95037-2845.